

Issued by Notified Body No. 0402 according to Regulation (EU) 2016/425 annex V, regarding:

Visor for protection against droplets and splashes of liquids

Issued to

GELA PLAST AB

Långgatan 6, SE-334 33 Anderstorp, Sweden

Product description and product name

Visor for protection against droplets and splashes of liquids. For details and limitations, see appendix.

Product name: GELA 225 VISIR

Technical documentation

The manufacturer's technical file, latest dated 2020-05-26.

Certificate

RISE Research Institutes of Sweden AB, Notified Body no. 0402, hereby certifies that the manufacturer's technical file and the product have been inspected in accordance with a simplified procedure based on annex V to regulation (EU) 2016/425 and found to fulfil the requirements. The product is judged to belong to category II according to annex I to the regulation. The EU Type-Examination is based on a type test in accordance with RISE Method MET5659.

Marking

The product shall be marked: "1 MET5659 (COVID-19) 3".

Miscellaneous

The manufacturer's information in Swedish on use, cleaning, storage etc. has been inspected and found to fulfil the relevant requirements for the intended use. This information must be available in a language understood where the product is to be sold.

Validity

The certificate remains valid as long as the conditions laid down in the specification in reference are not modified significantly or at the latest until 2020-12-31. The validity has been limited since the certification is based on a simplified procedure, and the products are ment to be used only for a limited period and a specified use.

Martin Tillander

Certificate C900105 | issue 1 | 2020-05-26

RISE Research Institutes of Sweden AB | Certification

Box 857, SE-50115 Borås, Sweden

+46 10 516 50 00 | certifiering@ri.se | www.ri.se

2P03386

This document is the property of RISE and may not be reproduced other than in full, except with the prior written approval by RISE



Product information




Photo of GELA 225 VISIR

Limitations of use:

The visor is intended to be used only during the Coronavirus disease (COVID-19) pandemic and for the purpose to protect the bearer from droplets and splashes of liquids originating from other persons. The visor is intended for re-use.

Material, frame: PC/ABS

Material, visor: PET-A, thickness 0.3 mm

Certificate history

Date	Issue	
2020-05-26	1	Certificate issued

Certificate C900105 | issue 1 | 2020-05-26

RISE Research Institutes of Sweden AB | Certification